Arabidopsis Metabolome Analysis Using Infusion ESI FT-ICR/MS

著者	Motohashi Reiko, Satou Masakazu, Myouga Fumiyoshi, Oikawa Akira, Ohta Daisaku
journal or publication title	Bio-protocol
volume	5
number	9
page range	e1463
year	2015-05-05
出版者	LLC
権利	(C) 2015 Bio-protocol LLC.
URL	http://hdl.handle.net/10297/9163

doi: 10.21769/bioprotoc.1463

-Category by field: Plant Science > Plant biochemistry > Other compound
Category by field: Plant Science > Plant metabolism > Metabolomics
Category by organism: Plants > Arabidopsis > Whole plant > Other compound

***Arabidopsis* Metabolome Analysis Using Infusion ESI FT-ICR/MS**

Reiko Motohashi^{1, 2*}, Masakazu Satou^{2, 3}, Fumiyoshi Myouga^{2, 4}, Akira Oikawa^{5, 6} and Daisaku Ohta⁵

¹Faculty of Agriculture, Shizuoka University, Shizuoka, Shizuoka 422-8529 Japan; ²Plant Science Center, RIKEN, Yokohama, Kanagawa, 230-0045 Japan (Present name: Center for Sustainable Resource Science (CSRS); ³Present address: Dragon Genomics Center, TAKARA BIO INC., Ootsu, Shiga 520-2198 Japan; ⁴Present address: RIKEN Center for Sustainable Resource Science, 1-7-22 Suehirocho, Tsurumi-ku, Yokohama, Kanagawa 230-0045, Japan; ⁵Graduate School of Agricultural and Biological Science, Osaka Prefecture University, Sakai, Osaka, 599-8531 Japan; ⁶Present address: Faculty of Agriculture, Yamagata University, Tsuruoka, Yamagata 997-8555 Japan

*For correspondence: armotoh@ipc.shizuoka.ac.jp

[Abstract] We made the method for Arabidopsis metabolome analysis based on direct-infusion Fourier transform ion cyclotron resonance mass spectrometry (FT-ICR/MS) (IonSpec). This method was sufficiently applied to metabolic phenotyping of Arabidopsis. This method is simple in that after homogenizing samples, powdered samples are dissolved in extraction solvents (acetone and methanol) to 20% fresh weight/volume. Extracted sample solutions are dried and dissolved in 50% (v/v) acetonitrile. Mass analysis using FT-ICR/MS (IonSpec) is performed in positive and negative ionization operation modes. Mass spectra are acquired over the 100-1000 m/z range and accumulated to improve the S/N ratio.

Materials and Reagents

1. 3-week-old Arabidopsis plants
2. Liquid nitrogen
3. Acetone, (HPLC grade) Wako 014-08681
4. Methanol, (HPLC grade) Wako 134-14523
5. Nitrogen gas
6. 50% (v/v) acetonitrile, (HPLC grade) Wako 018-19853, Use Distilled water to dilute acetonitrile to 50%(v/v).
7. Acetic acid, Wako 014-20063
8. 28~30% ammonia solution, WAKO 016-03146

9. Ultrapure water (recommended but not required)

Equipment

1. An IonSpec Explorer FT-ICR/MS equipped with a 7-tesla actively shielded superconducting magnet. (Ionspec, Sold by Agilent)
2. Glass vial and teflon cap, GL Sciences 1030-46716
3. Glass pipette
4. 0.45 µm filters (PTFE), ADVANTEC DISMIC 13HP045AN
5. Heat block

Procedure

A. Preparation of samples for FT-ICR/MS analysis

1. Homogenize samples with liquid nitrogen to create powder samples of whole plants using mortars and pestles.
2. Dissolve in extraction solvents to 20% fresh weight/volume. Two extraction solvents, 100% acetone and 100% methanol, are used to elute various polar compounds.
3. Filter extracted sample solutions through 0.45 µm filters (PTFE).
4. Transfer filtered sample solutions to vials.
5. Transfer 1 ml of sample solution to a separate vial. Place the sample solution into a draft chamber and apply nitrogen gas using EN1-16 (TAITEC 0076417-000). At 40 °C solvents will evaporate using a heat block.
6. Dissolve samples in 50% acetonitrile (0.1 ~ 1 mL depending on sample varieties) and store at -80 °C.
7. Dilute samples 1:14 prior to ESI analysis.

B. FT-ICR/MS analysis

8. Perform Mass analysis using an IonSpec Explorer FT-ICR/MS (Ionspec) in positive and negative ionization operation modes. Ions are generated from an ESI source with a fused silica needle of 0.005-inch i.d. (Oikawa et al. 2006). Samples are infused using a Harvard syringe pump mode 22 at a flow rate of 0.5 to 1.0 µL/min through a 100 µL Hamilton syringe. Set the potentials on the electrospray emitters to 3.0kV and -0.3kV for the positive and the negative electrosprays, respectively.
9. For positive mode, add 99.5% formic acid (HPLC grade, WAKO 063-04192) to extracted sample solutions (step 7) at a final concentration of 0.1% (v/v). Positive mode internal standards: Lidocaine (anaesthetic) m/z 235.18104 (WAKO 120-02691), Prochloraz (agricultural chemical) m/z 376.03863 (WAKO 164-25131), Reserpine (alkaloid sedative

drug) m/z 609.28121 (WAKO 184-00691), Bombesin (peptide) m/z 810.41479 (WAKO 339-40861).

10. For negative mode, add 28–30% ammonia solution (WAKO 016-03146) to extracted sample solutions at a final concentration of 0.1% (v/v).

Negative mode internal standards: 2,4-D (plant hormone) m/z 218.96157 (WAKO 040-18532), Ampicillin (antibiotic) m/z 348.10180 (WAKO 017-20531), CHAPS 8detergent) m/z 613.388865 (WAKO 341-04721), (GluNAc)₄ (Tetra-*N*-acetylchitotetraose) m/z 829.32023 (Tokyo Chemical Industry T2910).

11. Mass spectra are acquired over the 100–1,000 m/z range and accumulate to improve the S/N ratio. The time period for accumulation depends on the total ion concentration. Analyze peaks using the IonSpec Omega ver.8 software. Proofread m/z of each peak referencing the internal standards. Measure the product ion mass spectra of each sample three times.

12. When ion peaks are detected at least twice out of three successive spectral scans, they are subjected to further data processing as ion signals from actual analytes.

13. A total of four mass spectral peaks from two different extraction solvents (methanol and acetone) and two ionization operation modes (positive and negative) are aligned using our in-house Java program. (If you are interested in this in-house Java program, please contact Prof. Ohta.)

C. Normalization of data

Apply global normalization to data. To avoid zero division, missing values are filled with 10⁵ as a background signal of FT-ICR/MS. Peak intensities are transformed using a logarithmic scale with a factor of 10. Four data matrices are used to apply global normalization.

Global normalization calculation methods are as follows:

1. Average intensity is calculated by dividing the total signal by the number of detected peaks in each spectrum.
2. Average signal is calculated for all spectra in each elution and charge pair.
3. Normalization factor is calculated for each spectrum by dividing the average intensity for each spectrum by total average intensity.
4. Normalized intensity is calculated by multiplying the raw intensity in each spectrum by the previously calculated normalization factor. Empirical formulas are inferred by the accuracy of the FT-ICR/MS. Because the sample ions become adduct ions to attach protons and sodium ions, etc. on the ESI source, we assumed the following were involved in the detected peaks [M+H]⁺, [M+Na]⁺, [M+K]⁺, [M+H+methanol]⁺, [M+ammonium]⁺ in positive ionization mode and [M-H]⁻ in negative ionization mode.
5. Search for candidate compounds using KEGG (<http://www.genome.jp/kegg/>), NIST

(<http://webbook.nist.gov/chemistry/>) and KNApSack (Shinbo *et al.*, 2006).

Notes

1. In order to obtain reliably reproducible data, it is imperative that the plant growth environment be as uniform / consistent as possible. For example depending on light environment the value of a plant chlorophyll fluorescence will fluctuate. The amount of light a plant receives when next to the side light on the growth incubator is completely different from the light it receives when on the center of the shelf. For the reason it is important to shuffle the location of growth plates, etc.

Recipes

1. It is important that FT-ICRMS maintains a steady vacuum. We must also be careful about the size of the sample cone on the end of the ESI spray nozzle. It is also important to make minor adjustments where necessary to the direction of ESI ionspray, in an effort to capture as many ions into ICR cells as possible.
2. FT-ICRMS can have problems when the concentration of plant extraction samples is too high, or when other unwanted substances are included in samples. Regular baking and removal of ions from the chamber inner wall as well as cleaning of the ion source is required.

Acknowledgments

This work was supported by the Ministry of Education, Culture, Sports, Science and Technology (Japan) [Grants-in-Aid for Scientific Research (No.17681022 to R.M.)].

FT-ICR/MS (IonSpec) machine belongs to Osaka Prefecture University, and for this project it was used according to directions as indicated by Professors Ohta and Oikawa of the same university. Data normalization techniques were established by Mr. Satou.

This protocol is modified and appended referencing the original, as featured in Integrated analysis of transcriptome and metabolome of Arabidopsis albino or pale green mutants with disrupted nuclear-encoded chloroplast proteins. Satou M, Enoki H, Oikawa A, Ohta D, Saito K, Hachiya T, Sakakibara H, Kusano M, Fukushima A, Saito K, Kobayashi M, Nagata N, Myouga F, Shinozaki K, Motohashi R. Plant Mol Biol. 2014 Jul;85(4-5):411-28

References

1. Oikawa A, Nakamura Y, Ogura T, Kimura A, Suzuki H, Sakurai N, Shinbo Y, Shibata D, Kanaya S, Ohta D.(2006) [Clarification of pathway-specific inhibition by Fourier transform](#)

152 [ion cyclotron resonance/mass spectrometry-based metabolic phenotyping studies.](#) Plant
153 Physiol. 142(2):398-413.
154 2. Shinbo, Y., Nakamura, Y., Altaf-Ul-Amin, M., Asahi, H., Kurokawa, K., Arita, M., Saito, K.,
155 Ohta, D., Shibata, D. and Kanaya, S. (2006). [KNAPSAcK: a comprehensive](#)
156 [species-metabolite relationship database.](#) *Biotech Agric Forest* 57: 165-181.